



CERTIFICATE OF ANALYSIS

TSK-00001 SPLASH™ Kit: pTau-217 Singleplex

Quantity: 96 reactions

Lot Number: A123456

Expiration Date: 28-Mar-2027

Storage: -30 to -10°C (TSK-XXXXX-1)

2 to 8°C (TSK-XXXXX-2)

Note: For Research Use Only. Not for use in diagnostic procedures.

Component	Part Number	Lot Number
Sample Buffer	SPL-00013	A123456
Probe Buffer	SPL-00014	A123456
Wash Buffer	SPL-00015	A123456
Magnetic Beads	SPL-00016	A123456
Probe A	SPL-00009	A123456
Probe B	SPL-00010	A123456
Standard	SPL-00011	A123456
CheckPoint	SPL-00012	A123456
AmPLAmix	AMP-01050	A123456

QUALITY CONTROL

Parameter	Method	Requirement	Result
Functional QC	qPCR	dCq relative to blank > 5 Cq	Pass
pH of Sample Buffer	pH meter	pH 7.0 - 8.0	Pass
pH of Probe Buffer	pH meter	pH 7.0 - 8.0	Pass
pH of Wash Buffer	pH meter	pH 7.0 - 8.0	Pass

Quality authorized by :

Date: 1-Apr-2027

Taudia
1030 Brittan Ave., San Carlos, CA 94070
For inquiries, contact us at: info@taudia.com